

Remote electrode

Art. no. 13E400=*

Art. no.	Connection	Length	Frequency
13E400=G90-50	2-pin	90 mm	50 Hz
13E400=G90-60	2-pin	90 mm	60 Hz
13E400=G140-50	2-pin	140 mm	50 Hz
13E400=G140-60	2-pin	140 mm	60 Hz

ⓘ Notice: Six of the 13E400=* remote electrodes are required per prosthesis.

Remote electrode

Art. no. 13E401=*

Art. no.	Connection	Length	Frequency
13E401=G90-50	3-pin	90 mm	50 Hz
13E401=G90-60	3-pin	90 mm	60 Hz
13E401=G140-50	3-pin	140 mm	50 Hz
13E401=G140-60	3-pin	140 mm	60 Hz

ⓘ Notice: Two of the 13E401=* remote electrodes are required per prosthesis.