



Transtibial Silicon Liner

Patient Level: Active P/N 3022

- The AGILE liner(s) offer a full matrix which enhances control for higher activity and secure fit.
- Helps in stabilizing the limb, yet provides comfort
- · Suitable for moderate to high activity patients
- The enhanced distal matrix is part of the fabric covering of the liner and improves distal elongation control of the liner for both locking and cushion liner applications.
- A full matrix outer fabric covering that controls full length elongation and is available in both locking and cushion liner applications.
- Six-month warranty
- ADAPT and AGILE have the same thickness profile, thus are interchangeable if needed.

LOCKING P/N 3022L-(size)



CUSHION P/N 3022C-(size)



P/N	Description - LOCKING
3022L-18	AGILE Locking Liner, 18
3022L-20	AGILE Locking Liner, 20
3022L-22	AGILE Locking Liner, 22
3022L-23.5	AGILE Locking Liner, 23.5
3022L-25	AGILE Locking Liner, 25
3022L-26.5	AGILE Locking Liner, 26.5
3022L-28	AGILE Locking Liner, 28
3022L-30	AGILE Locking Liner, 30
3022L-32	AGILE Locking Liner, 32
3022L-34	AGILE Locking Liner, 34
3022L-36	AGILE Locking Liner, 36

P/N	Description - CUSHION
3022C-18	AGILE Cushion Liner, 18
3022C-20	AGILE Cushion Liner, 20
3022C-22	AGILE Cushion Liner, 22
3022C-23.5	AGILE Cushion Liner, 23.5
3022C-25	AGILE Cushion Liner, 25
3022C-26.5	AGILE Cushion Liner, 26.5
3022C-28	AGILE Cushion Liner, 28
3022C-30	AGILE Cushion Liner, 30
3022C-32	AGILE Cushion Liner, 32
3022C-34	AGILE Cushion Liner, 34
3022C-36	AGILE Cushion Liner, 36




2691 Saturn Street | Brea, CA 92821 Tel: (714) 524-0663 | Fax: (714) 364-8113



Rm. 911 43, Yangsan-ro Yeongdeungpo-gu, Seoul, Korea





Note: When placing an order, please use the below formula:

ADAPT TT Locking Liner (3012L)- indicate size

Example, if you were ordering a size 28, the part number would be: 3012L-28

AGILE TT Locking Liner (3022L)- indicate size


Example, if you were ordering a size 32, the part number would be: 3022L-32

HOW TO MEASURE Transtibial Liner

- 1. Measure the circumference of the residual limb at 4cm from the distal end
- 2. Choose the liner size, which corresponds to the measurement or one smaller.
- 3. Never choose a liner size, which is larger than the measurement



ST&G Liner Thickness Comparison





CORPORATE OFFICE

2691 Saturn Street | Brea, CA 92821 Tel: (714) 524-0663 | Fax: (714) 364-8113



Yeongdeungpo-gu, Seoul, Korea

