




PATIENT FITTING INSTRUCTIONS




Place the corset around your waist and center the brace on your back, over the spine. Slide your fingers into the hand grips of the lower elastic bands and hold the left flap firmly over your abdomen.




Pull the right side of the corset and attach it to lower left elastic band, overlapping the right side over the left.



Slide your fingers into the hand grips of the upper elastic straps and pull simultaneously away from your body and then towards each other as pictured above, securing the straps to the front of the brace.



Finally, ensure that the orthosis fit is snug and does not cause any discomfort.



Evotec TLSO w/ ATE Instructions: With the Evotec TLSO corset open and the anterior thoracic extension (ATE) attached to the left side only, slip your arm through the strap opening of the ATE so that the strap rests on your shoulder. Your caregiver should have already adjusted the ATE to the appropriate settings for your body. Follow steps 1-4 above to attach the corset and then snap the right side of the sternal panel into the buckle.

Evotec TLSO w/ DLK INSTRUCTIONS: With the Evotec TLSO corset open and the sternal panel attached to the left side only, slip your arms through the loops of the DLK as if putting on a backpack. Your caregiver should have already adjusted the DLK and the position of the left front floating panel specifically to the appropriate settings for your body. Follow steps 1-4 above to attach the corset then snap the right side of the sternal panel into the other side of the buckle and pull the shoulder straps of the DLK simultaneously to achieve the appropriate amount of tension as directed by your caregiver.

Care of Your Evotec

Remove plastic inserts from front and back panels. Hand wash at 86 F with mild soap. When drying, do not expose to direct heat (heater or sunlight); air dry. Never iron or dry clean. If not rinsed well after washing, residual soap may cause skin irritation and product deterioration. It is not uncommon for your body to change size under the spinal brace, because of the compression of the brace and because of a reduction in post-surgical swelling. If you change enough in size that you can no longer adjust the straps enough to get good compression, see your practitioner for a follow-up visit to resize or replace your brace. Your practitioner knows your medical status and is using the spinal brace as one part of your total care. It is important to follow his or her instructions exactly, even if they differ from some of the preceding guidelines. **WARRANTY:** Optec USA products are warranted for free repair or replacement due to manufacturer's defects within 90 days of the date purchased or the date fitted.

©Copyright ©2006 Velcro Industries B.V. All rights reserved. Velcro Industries B.V. VELCRO® and VELCRO with the FLYING-V® (DESIGN) are registered trademarks of Velcro Industries B.V.

OPTEC

DESIGN: LSO TLSO ATTACHMENTS: DLK PPK SPK PPX SPX

2S / XS (22 1/2 - 34 1/2) **SM / MD** (28 1/2 - 40 1/2) **LG / XL** (34 1/2 - 46 1/2)

2X / 3X (40 1/2 - 54 1/2) **4X / 5X** (46 1/2 - 60 1/2)