



# SUPER SLING


## Application Instructions



1. Place the arm in the sling with the shoulder strap detached and the front panel open. The elbow should be as far back into the sling as possible while securing the front panel strap through the lower D-ring.



2. Secure the shoulder strap diagonally over the unaffected shoulder and fasten through the middle D-ring so the end of the strap is at mid-chest level. Adjustments to the length of the shoulder strap can be made at the D-ring on the back end of the sling. For additional comfort, the thumb may be placed through the loop inside the sling.



3. Place the pillow at the waistline. With the large end facing forward, attach the sling to the pillow adhering adhesive strips. Fasten the waist strap through the D-ring on front end of pillow. The waist strap length can be adjusted at the D-ring on the back end of the pillow.

*Note: The pillow is designed to keep the arm in approximately 15° of abduction. For desired internal or external rotation, adjust the pillow by sliding it forward or back along the waistline.*

**WARNING:** This product is intended for application by licensed healthcare practitioners as directed by a physician or other medical authority. If you experience any pain, swelling, sensation changes or any unusual reactions while using this product, consult your physician immediately. Federal law restricts this device to sale by or on the order of a physician or duly licensed practitioner.

*Notice: This device is only one element in the overall treatment program administered by a medical professional. While every effort has been made to obtain the maximum compatibility of function, strength, durability and comfort, there is no guarantee that injury will be prevented through the use of this product.*