



## Sizing

### **Measure circumference of the hips at the widest point.**

Due to a wide range in waist and hip difference in many patients, size may need to be adjusted up or down by one size for a proper fit. Correct sizing is essential for patient comfort and product effectiveness. The product should fit tightly enough to provide even, constant pressure without having any impact on the patient's breathing. The product should not impinge on the patient's ribs or groin when seated.

Please Indicate Size when Ordering

### **Sizes Hip Measurement at Widest Point**

Small	30 – 45" (76.2 – 114.3 cm)
Large	46 – 62" (116.8 – 157.5 cm)
XLarge	46 – 78" (116.8 – 198.1 cm)