



With ankle in neutral position, measure circumference diagonally around calcaneous and malleoli.

Correct sizing is essential for patient comfort and product effectiveness. The product should fit tightly enough to provide even, constant pressure.

Please Specify Size when Ordering

Sizes	Measurements
Small	7" (17.8 cm)
Medium	8" (20.3 cm)
Large	9" (22.9 cm)
XLarge	10" (25.4 cm)