

Fabrication/Installation Instructions

1. Use of molding dummies or a *Tamarack hand-molding tool* (T-740-1 series) is necessary to generate a properly fitting cavity for the Tamarack Flexure Joints™.
2. Position each molding dummy* so the midpoint is located on/near the axis of joint motion (*Fig. A*) and fix in place (*Fig. B*). Additional plaster build-up may be necessary when you plan to use dorsiflexion assist flexures (742 series). They bulge inward during plantar flexion.
3. If a stockinette is pulled over the molding dummies before vacuum forming, it must be very thin/sheer. Any excess stockinette thickness will prevent a proper, snug fit of the cavity around the flexure.
4. When vacuum forming the thermoplastic or laminating, a small amount of extra material may be added at the Achilles Tendon area for a broader plantar flexion stop, if desired.
5. After cooling (or thermosetting), remove the plastic shell from the model, and extract the molding dummies. Using a thin-bladed saw (a fine-toothed coping saw is best), separate the foot section from the calf section (*Fig. C*). Do not use a cast saw for foot-calf separation (too much material is lost along a ragged, wide cut line). Sharp edges along the separation line should be beveled off with a hand deburring tool (*Fig. D*). Sanding or grinding will reduce flexure coverage and lessen the ability of the cavity to properly anchor and control the flexure.
6. a. Free Motion (740 series) – Grind a small “V” anterior to the midline of each cavity (*Fig. E*). Ensure that the “V” does not extend back past the center of the cavity.
 b. Dorsiflexion Assist (742 series) – Use a small radius tool to grind/form a “U” shaped anterior clearance (*Fig. F*). If plantar flexion range is required, remove material posteriorly as necessary (*Fig. G*) to provide needed clearance.
7. Use a *Tamarack hand punch tool* (T-740-2 series) to precisely locate and punch holes for the flexure anchoring screws (*Fig. H*). Large and medium flexures require 4.5mm (3/16 inch) diameter screw clearance hole; the pediatric size requires 4.0mm (5/32 inch) diameter.
8. Insert the Tamarack Flexure Joints™ and secure with the anchoring screws (Large and Medium, M4 x 9; Pediatric, M3.5 x 7). Depending upon the thickness of the plastic shell, it may be necessary to adjust the length of the screws. The screw must not protrude inside the AFO. Use a removable thread-locking compound on the screws. Longer screws are available on request.
9. A properly installed 740 series flexure will show no gapping along the separation cut except in the “V” or “U” shaped anterior clearance area.




Fig. A




Fig. B




Fig. C




Fig. D




Fig. E




Fig. F




Fig. G



Fig. H



Tamarack
HABILITATION TECHNOLOGIES, INC.

Flexure Joint™ Selection Guide

Standard Flexure Installation: Models 740 & 742

Using the standard molding cavity, any model of the same size can be interchanged for a variety of assist levels. See the chart on the opposite side for assist levels of several combinations.

Variable Assist (TVA) Installation: Model 743 (not shown)

Using the TVA molding cavity, the amount of assist can be adjusted with a simple hex driver. Assist level options are continuous from near zero to over 50% above standard.

For more information, call 1-866-795-0057 or visit www.oandp.com/tamarack

