

## Parts Included



Wave Suction Band



50ml Silicone Adhesive

### Wave Suction Band sizes

Part Number	Size	Circumference	Width	Thickness
CD130BM	Medium	23cm - 30cm	7.62cm	4mm
CD130BL	Large	30cm - 38cm	7.62cm	4mm
CD130BX	Extra Large	38cm - 48cm	7.62cm	4mm
CD130BXX	Extra Extra Large	48cm - 62cm	7.62cm	4mm

**Coyote**®

# CD130BL

## Large Wave Suction Band

Circumference 30cm - 38cm, Width 7cm, Thickness 5mm

### Fabrication Instructions



**Weight limit: 265 lbs.**

2-year warranty against manufacturer defects, excessive wear or breakage.

External Prosthetic Components



EC REP

Advena Limited  
Tower Business Centre  
2nd Flr, Tower Street  
Swatar, BKR 4013  
Malta

CE MD  
CD130BL.RevA.07312024

EN | Instructions for Use  
DE | Gebrauchsanweisung  
FR | Notice d'utilisation  
ES | Instrucciones para el uso  
IT | Istruzioni per l'uso  
NO | Bruksanvisning  
DA | Brugsanvisning  
SV | Bruksanvisning  
EL | Οδηγίες Χρήσης  
FI | Käyttöohjeet  
NL | Gebruiksaanwijzing  
PT | Instruções de Utilização

PL | Instrukcja użytkowania  
CS | Návod k použití  
TR | Kullanım Talimatları  
RU | Инструкция по использованию  
JA | 取扱説明書  
ZH | 中文说明书  
KO | 사용 설명서



[www.coyote.us/instructions-waveband](http://www.coyote.us/instructions-waveband)

**Coyote**

Manufactured by Coyote®  
419 N. Curtis Rd., Boise, Idaho 83706  
(208) 429-0026 | [www.coyote.us](http://www.coyote.us)



## Attaching Seal Band

Clean liner attachment area with acetone prior to gluing.




**1** With liner on patient position band in desired location.



**2** Mark band location on distal and proximal sides of band.



**3** For unconventional band position mark band in multiple locations.



**4** For unconventional band position mark band in multiple locations.

**5** Wrap the liner with electrical tape or vinyl tape next to the band without over lapping it.




**6** Tape off both ends of the band.




**7** Remove the band



**8** Add extra tape to protect a wider area.



**9**



**10** Put a bead of silicone adhesive around the liner about the thickness of a pencil.



**11** Smooth the silicone out with a tongue depressor. Make sure to saturate the material.




**12** Check the edges for areas that need more silicone adhesive.




**13** Smooth the silicone out.




**14** Carefully place the seal band over the glued area.




**15** Center the band with tape to the edges.




**16** Push the air bubbles out from under the band




**17** Remove the tape from the liner.



**18** Create a seal band edge with silicone adhesive.




**19** Roll tape next to band edge approximately 1/4 inch space from band.



**20** Smooth silicone into fabric and band. Making a sealed edge.



**21** Remove tape.



**22** Check your work and let the silicone set up for at least 8 hours before putting it on patient.

#### Wave Suction Band sizes

Part Number	Size	Circumference	Width	Thickness
CD130BM	Medium	23cm - 30cm	7.62cm	4mm
CD130BL	Large	30cm - 38cm	7.62cm	4mm
CD130BX	Extra Large	38cm - 48cm	7.62cm	4mm
CD130BXX	Extra Extra Large	48cm - 62cm	7.62cm	4mm

Lay-up instructions are helpful hints on how to work with the lock and connector. Actual lay-ups are responsibility of the technician and/or practitioner.